

	OPERASI PERKHIDMATAN SOKONGAN PUSAT KESIHATAN UNIVERSITI Kod Dokumen: PKU/BR07/LPT
	BORANG UJIAN PERALATAN PERLINDUNGAN SINARAN (LEAD PROTECTION TEST)

Perkhidmatan/Unit :
Tarikh Pengujian :
Tarikh Pengujian Semula :

No	Lead Protection Accessories /ID	Measured Readings in diameter (cm)	Calculation	Limit: Sum or aggregate area of holes or crack < 10cm ² , < 0.2 cm ² in the reproductive region, and <1 cm ² int the thyroid region
1.				Pass / Fail Leakage area: Region :
2.				Pass / Fail Leakage area: Region :
3.				Pass / Fail Leakage area: Region :
4.				Pass / Fail Leakage area: Region :
5.				Pass / Fail Leakage area: Region :
6.				Pass / Fail Leakage area: Region :

Peralatan Perlindungan Sinaran berplumbum layak dilupuskan sekiranya hasil pemeriksaan mempunyai kriteria berikut:

- jika jumlah kawasan lubang atau retakan di gaun plumbum melebihi 10 cm² atau;
- jika mana-mana lubang atau retakan berada di kawasan pembiakan (reproductive region) tidak melebihi 0.2 cm² atau;
- jika thyroid shield mempunyai lubang atau retakan di bahagian tiroid melebihi 1cm².

Pengujian Dilakukan Oleh:

.....
(Juru Xray)
Pusat Kesihatan Universiti

Pengesahan Penerimaan Oleh:

.....
(Ketua Perkhidmatan)
Pusat Kesihatan Universiti